

Australian/New Zealand Standard™

Stationary source emissions

**Part 3: Determination of odour
concentration by dynamic olfactometry**



AS/NZS 4323.3:2001

This Joint Australian/New Zealand Standard was prepared by Joint Technical Committee EV-007, Methods for Examination of Air. It was approved on behalf of the Council of Standards Australia on 31 July 2001 and on behalf of the Council of Standards New Zealand on 1 August 2001. It was published on 12 September 2001.

The following interests are represented on Committee EV-007:

Australian Aluminium Council
Australian Chamber of Commerce and Industry
The Australian Gas Association
Australian Industry Group
Bureau of Meteorology, Australia
Clean Air Society of Australia and New Zealand
CSIRO Energy Technology
Department of Environment, Heritage and Aboriginal Affairs, S.A.
Department of Environmental Protection, W.A.
Electricity Supply Association of Australia
Environment Australia
Environment Protection Authority, N.S.W.
Environment Protection Authority of Victoria
Environment Protection Agency (Queensland)
National Association of Testing Authorities, Australia
National Health and Medical Research Council, Australia
Works Environmental Management, New Zealand

Additional interest:

Independent test laboratories

Keeping Standards up-to-date

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about joint Australian/New Zealand Standards can be found by visiting the Standards Australia web site at www.standards.com.au or Standards New Zealand web site at www.standards.co.nz and looking up the relevant Standard in the on-line catalogue.

Alternatively, both organizations publish an annual printed Catalogue with full details of all current Standards. For more frequent listings or notification of revisions, amendments and withdrawals, Standards Australia and Standards New Zealand offer a number of update options. For information about these services, users should contact their respective national Standards organization.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Please address your comments to the Chief Executive of either Standards Australia International or Standards New Zealand at the address shown on the back cover.

Australian/New Zealand Standard™

Stationary source emissions

Part 3: Determination of odour concentration by dynamic olfactometry

First published as AS/NZS 4323.3:2001.

COPYRIGHT

© Standards Australia/Standards New Zealand

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Jointly published by Standards Australia International Ltd, GPO Box 5420, Sydney, NSW 2001 and Standards New Zealand, Private Bag 2439, Wellington 6020

ISBN 0 7337 4087 1

PREFACE

This Standard was prepared by the Joint Standards Australia/Standards New Zealand Committee EV-007, Methods for Examination of Air.

The Standard is based on a CEN pre-draft of the same title, produced by CEN/TC 264, Air Quality, Working Group 2, Odours. The quality criteria were validated in an interlaboratory comparison for olfactometry in 1996.

Area source sampling is not covered in this Standard as it is considered such an important area, that it warrants a future Standard in its own right. AS 4323.1, *Stationary source emissions, Part 1: Selection of sampling positions* should be referenced when sampling for odour from stacks or flues.

The objective of this Standard is to provide a method for the determination of the odour concentration of a gaseous sample using dynamic olfactometry with a panel of human assessors being the sensors.

The term 'informative' has been used in this Standard to define the application of the appendix to which it applies. An 'informative' appendix is only for information and guidance.

CONTENTS

	<i>Page</i>
1 SCOPE	4
2 APPLICATION	4
3 REFERENCED DOCUMENTS.....	5
4 DEFINITIONS.....	5
5 NOTATION.....	12
6 THE UNIT OF MEASUREMENT	14
7 PRINCIPLE OF MEASUREMENT.....	16
8 PERFORMANCE QUALITY REQUIREMENTS	16
9 MATERIALS, GASES AND PANEL MEMBERS	24
10 SAMPLING.....	30
11 PRESENTATION OF ODORANTS TO ASSESSORS	33
12 DATA RECORDING, CALCULATION AND REPORTING.....	35
 APPENDICES	
A PHYSIOLOGICAL PRINCIPLES.....	41
B EXAMPLE OF ACCURACY CALCULATION FOR ODOUR MEASUREMENTS WITHIN ONE LABORATORY	44
C EXAMPLE OF CALCULATION OF INSTRUMENTAL ACCURACY AND INSTABILITY.....	46
D EXAMPLE OF CALCULATIONS FOR PANEL SELECTION.....	48
E EXAMPLE OF THE CALCULATION USED TO DETERMINE THE NUMBER OF ODOUR CONCENTRATION MEASUREMENTS REQUIRED TO ACHIEVE A DEFINED PRECISION.....	49
F EXAMPLE OF THE CALCULATION OF THE ODOUR CONCENTRATION FROM A SET OF PANEL MEMBER RESPONSES	51
G EXAMPLE OF THE CALCULATION OF THE ODORANT FLOW RATE (STANDARD CONDITIONS) FOR A WET EMISSION	55
H EXAMPLE OF THE CALCULATION USED TO DETERMINE THE NUMBER OF ODOUR CONCENTRATION MEASUREMENTS REQUIRED TO DETECT A DIFFERENCE BETWEEN TWO MEANS.....	56

STANDARDS AUSTRALIA/STANDARDS NEW ZEALAND

Australian/New Zealand Standard
Stationary source emissions**Part 3: Determination of odour concentration by dynamic olfactometry****1 SCOPE**

This Standard specifies a method for the measurement of odour concentration of pure substances, defined and undefined mixtures of gaseous odorants in air or nitrogen, using dynamic olfactometry with a panel of human assessors being the sensors. The unit of measurement is the odour unit ou. The odour concentration is measured by determining the dilution factor required to reach the detection threshold. The odour concentration at the detection threshold is by definition 1 ou. The odour concentration is then expressed in terms of multiples of the detection threshold. The range of measurement is typically from 10^1 to 10^7 ou (including pre-dilution).

The relation between emissions, dispersion, exposure and annoyance is not within the scope of this Standard.

2 APPLICATION**2.1 General**

The primary application of this Standard is to provide a common basis for evaluation of odorant emissions.

2.2 Inclusions

The characterization of odour emissions requires detailed measurement of the gas velocity that shall be performed according to the relevant Standards included in the normative references (see Clause 3).

The field of application includes the following:

- (a) The measurement of the mass concentration at the detection threshold of pure odorous substances in grams per cubic metre.
- (b) The measurement of the odour concentration of mixtures of odorants in odour units (ou).
- (c) The measurement of the emission rate of odorous emissions from point sources and surface sources (with and without an outward flow), including pre-dilution during sampling.
- (d) The sampling of odorants from emissions of high humidity and temperature (up to 200°C).
- (e) The determination of effectiveness of end-of-pipe devices used to reduce odour emissions.

2.3 Exclusions

The field of application does not include:

- (a) The measurement of odours potentially released by particles of odorous solids or droplets of odorous fluids suspended in emissions.



SAI GLOBAL

This is a free 6 page sample. Access the full version online.

The remainder of this document
is available for purchase online at

www.saiglobal.com/shop

SAI Global also carries a wide range of publications from a wide variety of Standards Publishers:



Click on the logos to search the database online.